

## **MLT118F GSR Finger Electrodes**

### *Transducer Series*

Bipolar finger electrodes for measuring galvanic skin response. Suitable for connection to the FE116 GSR Amp. MLT118F is MR unsafe, see MLT117F for use in MR environments.



#### **Caution**

Read “Statement of Intended Use” on our website or in “Getting Started with PowerLab” before use.

#### **Ordering Information:**

MLT118F GSR Finger Electrodes

For use with:

FE116 GSR Amplifier